

CERTIFICATE OF ANALYSIS

Trypsin 2,5% in HBSS

w/o Calcium, Magnesium and Phenol Red

Cat no.: ECB3051D

Batch no.: EUM02VQ

Expiry Date: 09/01/2028

PHYSICAL AND CHEMICAL ANALYSIS

	SPECIFICATIONS	RESULTS	
pH	7,3 ± 0,3	7,20	
OSMOLALITY ^[1]	390 ± 10	360	mOsm/kg
ACTIVITY	Pass	Pass	

STERILITY TESTING

	SPECIFICATIONS	RESULTS
BACTERIA and FUNGI ^[2]	NOT DETECTED	NOT DETECTED

REFERENCES

- ^[1] Freezing point EU Ph. 2.2.35
^[2] Culture

THIS CELL CULTURE REAGENT IS NOT ANIMAL COMPONENT FREE

This document has been produced electronically and is valid without a signature

Francesca Carlino
Quality Assurance Dept.

Date: 28/07/2026

ECB3051_batch_EUM02VQ_rev0

Euroclone SpA

Via Figino, 20/22 • 20016 Pero (MI) Italy • Ph. +39.02.38195.1 • Fax +39.02.33913713 • info@euroclone.it • www.euroclone.it
Sede legale: Viale Liguria, 40 • 20143 Milano (Italy)

C. F. P.IVA e n° Registro Imprese di Milano 08126390155 • REA Milano 1208538 Cap. Soc. Euro 230.000 € int. vers
Quality Management Systems and Enviromental certified according to EN ISO 9001, ISO 13485 and EN ISO 14001