

**CERTIFICATE OF ANALYSIS
HANK'S BSS
W/O CALCIUM W/O MAGNESIUM**

Cat N.: ECB4007L

Batch N.: EUM02KW
Expiry Date: 11/03/2029

PHYSICAL AND CHEMICAL ANALYSIS

	SPECIFICATIONS	RESULTS
pH	7.3 ± 0.3	7.50
OSMOLALITY ^[1]	280 ± 10%	276 mOsm/Kg
ENDOTOXIN ^[2]	< 1	< 0.005 EU/ml

STERILITY TESTING

	SPECIFICATIONS	RESULTS
BACTERIA AND FUNGI ^[3]	NOT DETECTED	NOT DETECTED

REFERENCES

- ^[1] Freezing-Point EU Ph. 2.2.35
- ^[2] Chromokinetic test Method D EU Ph. 2.6.14
- ^[3] Culture

THIS CELL CULTURE REAGENT IS ANIMAL COMPONENT FREE.

This document has been produced electronically and is valid without a signature

Francesca Carlino
Quality Assurance Dept.

DATE: 30/06/2025

ECB4007_batch_EUM02KW_rev0

Page 1 of 1