

CERTIFICATE OF ANALYSIS
TRYPSIN EDTA 1X IN PBS
W/O CALCIUM W/O MAGNESIUM W/O PHENOL RED

Cat N.: ECB3052D

Batch N.: EUM02OS
Expiry Date: 12/06/2027

PHYSICAL AND CHEMICAL ANALYSIS

	SPECIFICATIONS	RESULTS
pH	7.3 ± 0.3	7.36
OSMOLALITY ^[1]	290 ± 10%	283 mOsm/Kg
ACTIVITY	Cells detachment	PASS

STERILITY TESTING

	SPECIFICATIONS	RESULTS
BACTERIA AND FUNGI ^[2]	NOT DETECTED	NOT DETECTED

REFERENCES

- [1] Freezing-Point EU Ph 2.2.35
[2] Culture

THIS CELL CULTURE REAGENT IS NOT ANIMAL COMPONENT FREE.

This document has been produced electronically and is valid without a signature

Francesca Carlino
Quality Assurance Dept.

Date 02/10/2025