

**CERTIFICATE OF ANALYSIS
COLCEMID 10 ug/ml IN PBS**

Cat N.: EK0010

Batch N.: EUK02NR
Expiry Date: 14/05/2027

QUALITY PROFILE

	SPECIFICATION	RESULTS
pH	6.9 – 7.5	7.40
OSMOLALITY	265 - 325	284 mOsm/Kg
ENDOTOXIN	<10.000	<0.500 EU/ml

STERILITY PROFILE

	SPECIFICATIONS	RESULTS
BACTERIA AND FUNGI ^[3]	NOT DETECTED	NOT DETECTED

This document has been produced electronically and is valid without a signature

Francesca Carlino
Quality Assurance Dept.

DATE: 29/07/2025

EK0010_batch_EUK02NR_rev0

Page 1 of 1