

CERTIFICATE OF ANALYSIS

Dulbecco's PBS w/o Ca & Mg

Cat no.: ECB4004L

Batch no.: EUM02ZT

Expiry Date: 13/04/2030

PHYSICAL AND CHEMICAL ANALYSIS

	SPECIFICATIONS	RESULTS	
pH	7,5 ± 0,3	7,53	
OSMOLALITY ^[1]	290 ± 10%	276	mOsm/kg
ENDOTOXIN ^[2]	< 1,000	< 0,005	EU/ml

STERILITY TESTING

	SPECIFICATIONS	RESULTS
BACTERIA and FUNGI ^[3]	NOT DETECTED	NOT DETECTED

REFERENCES

- ^[1] Freezing point EU Ph. 2.2.35
- ^[2] Chromokinetic test Method D of EU Ph. 2.6.14
- ^[3] Culture

THIS CELL CULTURE REAGENT IS ANIMAL COMPONENT FREE.

This document has been produced electronically and is valid without a signature

Francesca Carlino
Quality Assurance Dept.

Date: 01/07/2026

ECB4004_batch_EUM02ZT_rev0

Euroclone SpA

Via Figino, 20/22 • 20016 Pero (MI) Italy • Ph. +39.02.38195.1 • Fax +39.02.33913713 • info@euroclone.it • www.euroclone.it
Sede legale: Viale Liguria, 40 • 20143 Milano (Italy)

C. F. P.IVA e n° Registro Imprese di Milano 08126390155 • REA Milano 1208538 Cap. Soc. Euro 230.000 € int. vers
Quality Management Systems and Enviromental certified according to EN ISO 9001, ISO 13485 and EN ISO 14001