

CERTIFICATE OF ANALYSIS

RPMI 1640 w/ stable Glutamine

Cat no.: ECM2001L**Batch no.:** EUM02T3**Expiry Date:** 08/10/2027

PHYSICAL AND CHEMICAL ANALYSIS

	SPECIFICATIONS	RESULTS	
pH	7.3 ± 0.3	7,29	
OSMOLALITY ^[1]	278 ± 10%	278	mOsm/kg
ENDOTOXIN ^[2]	< 1	< 0,005	EU/ml

STERILITY TESTING

	SPECIFICATIONS	RESULTS
BACTERIA and FUNGI ^[3]	NOT DETECTED	NOT DETECTED

CELL CULTURE TEST

CELL LINE	SP2/O-Ag14	
	SPECIFICATIONS	RESULTS
CELL GROWTH	75% of our reference	PASS

REFERENCES

- ^[1] Freezing point EU Ph. 2.2.35
- ^[2] Chromokinetic test Method D of EU Ph. 2.6.14
- ^[3] Culture

THIS CELL CULTURE MEDIA IS ANIMAL COMPONENT FREE

This document has been produced electronically and is valid without a signature

Francesca Carlino
Quality Assurance Dept.

Date: 16/04/2026

ECM2001_batch_EUM02T3_rev0

Euroclone SpAVia Figino, 20/22 • 20016 Pero (MI) Italy • Ph. +39.02.38195.1 • Fax +39.02.33913713 • info@euroclone.it • www.euroclone.it
Sede legale: Viale Liguria, 40 • 20143 Milano (Italy)C. F. P.IVA e n° Registro Imprese di Milano 08126390155 • REA Milano 1208538 Cap. Soc. Euro 230.000 € int. vers
Quality Management Systems and Enviromental certified according to EN ISO 9001, ISO 13485 and EN ISO 14001